

- All Communications to be addressed  
to the Director, NEIST  
With Attn: *Stores & purchase officer*

Telefax: 91-376-2372921

Email: [spopur@rrljorhat.res.in](mailto:spopur@rrljorhat.res.in)

**NORTH-EAST INSTITUTE OF SCIENCE AND TECHNOLOGY**  
**(Formerly Regional Research Laboratory)**  
**(Council of Scientific & Industrial Research)**  
**JORHAT-785 006, ASSAM, INDIA**

**By Speed Post**

File No. 3(POQ)56/10-11/PUR/T-96

Date: 02.02.2011.

Note: Please send your quotation in  
sealed cover **on or before 02.03.11**  
**up to 2.30 P.M.** and opening on  
**02.03.11 at 3.00 P.M.**

Dear Sirs,

Kindly arrange to send your offer in a SEALED COVER immediately quoting this office file reference to the terms & conditions overleaf.

Sl. No.	Description of Stores	Quantity
1	Microscope	2 Nos
2	Fluorescent Microscope	1 No

**DETAILED SPECIFICATIONS AS ENCLOSED SHEET**

Thanking you,

Yours faithfully,

**STORES & PURCHASE OFFICER**

## TERMS & CONDITIONS FOR TENDERS

1. We are prepared to consider offer of indigenous manufacture or of foreign make available from ready stock. Any offer to supply on forward Delivery Basis under supplier's own import quote license will also be considered. In such a case indicate ICT No. Name and address of foreign principals.
2. The price quoted should be as indicated as per one of the clauses given below. F.O.R destination by passenger/goods Train/Lorry service. FO.R dispatching station (name of station.) plus estimated freight charges free delivery at our Stores of the NEIST, JORHAT.
3. Prices are required to be quoted in units indicated in the inquiry, when quotations are given in terms of other units relationship between two sets of units should be furnished. Quantity discount's if any should also be indicated.
4. Your rates should include packing and forwarding charges in case of F.O.R station of despatch. Insurance should be arranged for the goods against all risks on warehouse-to-warehouse basis (Local suppliers should quote for delivery at the stores of the NEIST, JORHAT).
5. Your offer should be valid for sixty days from the date of opening of the Quotations. No revision in price will be allowed after accepting the quotation.
6. The full rate of Sales Tax should be clearly indicated wherever chargeable. This office is not eligible to issue "C" or "D" forms.
7. This office is exempted from paying excise duty (Notification No. 10/97 refers). Hence please quote your rates without Excise duty. Necessary Excise duty exemption certificate will be sent along with the order.
8. Specific mention should be made about the delivery period required. THE DELIVERY TIME WILL HAVE TO BE STRICTLY ADHERED TO INCASE AN ORDER IS PLACED against your offer.
9. In case any of the items required are available on DGS&D rate contract then please quote D.G.S & D rate, the contract references and also send a copy of the RATE CONTRACT.
10. The quotation should be duly signed by an authorized person of the firm. The name and status of the signatory should be clearly indicated. UNSIGNED QUOTATIONS WILL NOT BE CONSIDERED.
11. The quotations are to be enclosed in double cover. The inner cover should be SEALED WITH WAX AND SUPERSCRIBED as per subject mentioned above. The outer cover should bear enquiry no. and date, last date of submission and due date and time of opening the tender. Quotation delivered personally should be put in the tender box kept in the Main Gate (Gate No. 2) of NEIST, JORHAT.
12. Incase of equipment kindly let us know the organizations to whom you have supplied the same in the recent past.
13. Payment: Unless otherwise agreed payments will be made directly to the supplier(s) by e-payment [Please provide all bank details like Account Number, Branch Code Number etc. for E-PAYMENT] within 30 days receipt of the stores in good conditions and bill in order
14. The right of acceptance of the quotation will rest with the Director, NEIST, JORHAT who does not bind himself to accept the lowest quotation and reserves the right to reject or partially accept any or all the quotations received, without assigning any reason.
15. Tender containing erasure or alterations should be countersigned.
16. You shall indicate in the quotation the name of your Bankers and permanent Sales-Tax Account Number.
17. All disputes arising out of this shall be referred to the sole arbitration of the Director, NEIST, JORHAT and his award shall be final and binding on the parties to the dispute. The venue of arbitration shall be JORHAT. All suits in respect of this shall lie in courts of JORHAT jurisdiction.
18. Tender's conditions (Printed on the reverse), if any, or otherwise sent along with the tender shall not be binding on us.
19. Conditional tenders would be rejected outright.
20. The quotations are liable to be rejected if any of the above conditions are not complied with.

**STORES & PURCHASE OFFICER**



Contd page 2

**SPECIFICATION OF FLUORESCENT MICROSCOPE WITH PHOTOGRAPHIC ARRANGEMENTS :**

1 No

1. The Microscope should be suitable for BF, DF, Phase, PlasDIC, Fluorescence.
2. Transmitted light halogen illumination 50 watt for observation and photomicrography.
3. Hextuple nosepiece with objective 5x, 10x, 40x and 100x for fluorescence and bright field.
4. Large circular disc type intensity regulator arrange on both side of the stand should be placed just below the focus knob.
5. 6 – position filter wheel at the front equipped with the integrated shutter additional filter slider at the back.
6. Binocular tube minimum with 23 mm field of view.
7. Option for increasing the height of the microscope, continuously adjustable for specimen thickness from 0 to 110mm.
8. Should be completed with dust cover and dust protection arrangement from manufacturer.
9. Fluorescence filter should be with light trap system. Filter should be insertable and removable with push & click method.
10. Fluorescence filter (1) excitation 450 – 490, Dichromatic Mirror 510, Emission Filter 515 (3) excitation BP 546/12, beam splitter FT 580, emission filter LP 590.
11. Digital camera attachment with high quality minimum 2/3” chip size, Resolution 1360 x 1024 Pixel in 12 bit per channel CCD camera with Pixel Pitch : 6.45 x 6.45  $\mu\text{m}$ ,  
~ Gain : analog 1 to 8 linear, digital 1 to 10 linear-Hardware-Trigger connection (input and output).
12. Suitable computer with Software :
  - a) Multi-Fluorescence mode of up to 5 color filters
  - b) Programmable resolution,
  - c) Various measurements tools,
  - d) Reproducible data and Annotations
  - e) Storage as movie file Focus tool,
  - f) Multi-Focus Capture of images.

  
Stores & Purchase Officer