



**NORTH – EAST INSTITUTE OF SCIENCE AND TECHNOLOGY**  
**(Formerly Regional Research Laboratory)**  
**(Council of Scientific & Industrial Research)**

**Sub: CORRIGENDUM**

**dtd 3.2.2010**

The following amendment may please be incorporated in our earlier N I T No. II/2009-10 dated 09.01.10

Sl. No. 44 & Page 53 Weighing balance (Digital – High precision)

FOR	Readability: 0.001mg, Repeatability: 0.001 mg, Linearity: $\pm 0.002$ mg
READ	Readability: 0.001g, Repeatability: 0.001g, Linearity: $\pm 0.002$ g

The amendment is issued without prejudice with the other terms & conditions of our earlier N I T.

Sd/-  
Stores & Purchase Officer